## Which Chamber is Right for Me?

One size does not fit all when it comes to chambers. Use the chart below to find the most appropriate chamber.



HILD ADUL

## Tip: Use the Flow-Vu\* Inhalation Indicator

The *Flow-Vu*\* Indicator moves with inspiration and helps to provide assurance of correct use for medication delivery.

- · Helps coordinate actuation with inhalation
- Ensures proper seal
- Enables counting of breaths

## **Chamber Care and Replacement**



## 3 reasons to replace

- 1 Chambers should be replaced every 12 months for optimum medication delivery. AeroChamber Plus\* Flow-Vu\* chamber undergoes lifecycle testing for one year of regular use.
- 2 Transition to a new size.
  - Ensure children under 5 years of age are using the correct mask size
  - Children should transition to a mouthpiece around 5 years of age
- It is important to follow recommended cleaning guidelines. Not doing this can result in medication build-up in the valves which may impair performance.

For optimal performance chambers should be replaced every 12 months



